



EU Declaration of Conformity

Quantec Dementia Care Devices

Product Model Designation(s) & Description(s)

- QT631K Quantec Dementia Care Room Status Controller, Keyswitch Reset/Isolate
- QT631M Quantec Dementia Care Room Status Controller, Magnetic Reset/Isolate
- QT636 Quantec Interface Unit c/w Remote Socket
- QT637 Quantec Enuresis/General purpose Interface Socket
- QT645 Quantec PIR Power Interface c/w Isolating Keyswitch

Manufactured by C-TEC at:

Computationics Ltd (Trading as C-TEC) Challenge Way, Martland Park
Wigan, WN5 0LD, United Kingdom

are in conformity with the provisions of the following EU Directive(s):

- 2014/30/EU: Electromagnetic Compatibility Directive (EMC)
- 2014/35/EU: Low Voltage Directive (LVD)
- 2011/65/EU: Restriction of Use of certain Hazardous Substances Directive (RoHS)

The products to which this declaration relates have been tested to and are in conformity with the following standards:

- EN 50130-4:2011+A1:2014 Alarm Systems – Part 4: Electromagnetic compatibility - Product family standard: Immunity requirements for components of fire, intruder and social alarm systems.
- EN 61000-6-3: 2007+A1:2011 Electromagnetic compatibility (EMC) – Part 6-3: Generic standards - Emission standard for residential, commercial and light-industrial environments.
- EN50581: 2012 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

Empowered Signatory of Company

I, the undersigned, hereby declare on behalf of Computationics Limited that the products listed above comply with the requirements of the standards listed above which gives presumption of conformity with the essential requirements of the directives listed above.

Name: Stephen A. Stringfellow

Position: Research & Development Manager

Signature:

Date: 16.8.16